


Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: A1494 ACETONE REAGENT ACS			
Test	Target/UOM	Range	Result
ASSAY OF MATERIAL	>= 99.5%		99.8%
SOLUBILITY	PASSES TEST SOLUBILITY		PASSES TEST
TITRABLE ACID	MAX 0.0003 MEQ/G T ACID		0.00020 MEQ/G
TITRABLE BASE	MAX 0.0006 MEQ/G T BASE		< 0.0001 MEQ/G
ALDEHYDE (AS HCHO)	MAX 0.002% ALDEHYDE		< 0.002%
SUBSTANCES REDUCING PERMANGANATE	PASSES TEST REDUCE		PASSES TEST
%WATER	MAX 0.5% %WATER		0.2%
ISOPROPYL ALCOHOL	MAX 0.05% ISOPROPYL		< 0.005%
COLOR (APHA)	MAX 10		5
RESIDUE AFTER EVAPORATION	MAX 0.001%		< 0.0001%
Lot # 430102			
Made			



Quality Manager
Aqua Solutions, Inc.