



Certificate of Analysis

Certipur[®] Certified Reference Material

Producer:	Merck KGaA, Frankfurter Str. 250, 64293 Darmstadt, Germany
Accreditation:	The laboratory of Merck KGaA for reference materials in titrimetry is accredited as calibration laboratory according to DIN EN ISO/IEC 17025 .
Description of CRM:	Potassium iodate
Ordering number:	1.02404.0100
Lot number:	132404L
Composition:	Potassium iodate
Certified value and uncertainty:	99.76 % ± 0.05 % Mass fraction in weight percent with expanded measurement uncertainty The expanded measurement uncertainty U is calculated as $U = k \cdot u_{\text{characterisation}}$, where $k = 2$ is the coverage factor for a 95% coverage probability and $u_{\text{characterisation}}$ is the combined measurement uncertainty in accordance to DIN EN ISO/IEC 17025.
Method of analysis:	The certified mass fraction was determined by potentiometric titration with sodium thiosulfate as titration solution. The certified value is based on a molecular mass $M = 214.001$ g/mol dried substance.
Traceability:	This volumetric standard is directly traceable to the corresponding primary standard NIST SRM [®] potassium dichromate 136f. NIST: National Institute of Standards and Technology, Gaithersburg, USA.
Storage:	Store at +15 °C to +25 °C tightly closed in the original container and protect from light and moisture.
Application and correct use:	This volumetric standard is intended for standardisation of volumetric solutions in accordance to the chapter reagents of the Pharmacopoeia (Ph. Eur., USP). The volumetric standard potassium iodate must be dried at 130 °C for 2 hours before use.
Date of release:	2013/08/20
Minimum shelf life:	2018/08/31

Dr. Stefan Frey
(Laboratory Manager)