



Certificate of Analysis

1.09060.4000 Hydrochloric acid $c(\text{HCl}) = 0.1 \text{ mol/l}$ (0.1 N) TitriPUR®

Batch HC395640

Volumetric solution in accordance to the chapter reagents of the Pharmacopoeia (Ph Eur, USP).

Batch Values

Amount-of-substance concentration 0.1000 mol/l

The concentration of this volumetric solution was determined with volumetric standard Tris(hydroxymethyl)aminomethane (Merck article number 1.02408). The determined titer at 20°C was 1.000 with an expanded measurement uncertainty of ± 0.003 ($k=2$ coverage factor for 95% coverage probability). The certified value is traceable to primary standard NIST SRM 723e (NIST: National Institute of Standards and Technology, USA) by means of volumetric standard Tris(hydroxymethyl)aminomethane, measured in the accredited calibration laboratory of Merck KGaA in accordance to DIN EN ISO/IEC 17025.

Date of release (DD.MM.YYYY) 07.08.2013

Minimum shelf life (DD.MM.YYYY) 31.08.2016

Ayfer Yildirim

Responsible laboratory manager quality control

This document has been produced electronically and is valid without a signature.