

Certificate Of Analysis



Date of Release: 08.18.2011
Name: Buffer Solution, pH 7
Buffer pH 7.00 Concentrate
NIST Traceable
Item No: BX1627 All size codes
Lot / Batch No: 1220
Country of Origin: USA

Characteristic	Requirement		Results	Units
	Min.	Max.		
Appearance			Passes test	
Expiration Date			31-AUG-13	
pH @ 25°C	6.99	7.01	7.00	pH
SRM Disodium Hydrogen Phosphate			1861lg	
SRM Potassium Dihydrogen Orthophosphate			1861g	
SRM Potassium Hydrogen Phthalate (pH)			185h	

Joe Schoellkopf

Quality Control Manager

This document has been produced electronically and is valid without a signature.

EMD Chemicals Inc.

480 South Democrat Road
Gibbstown, NJ 08027

Phone: 856-423-6300
Fax: 856-423-4389
www.emdchemicals.com