

Certificate Of Analysis



Date of Release: 10.12.2011

Name: Buffer Solution pH 4.0
Buffer pH 4.00 pH $\hat{A}\pm$ 0.01, Red
NIST Traceable

Item No: BX1628, all size codes

Lot / Batch No: 1285

Country of Origin: USA

Characteristic	Requirement		Results	Units
	Min.	Max.		
Appearance			Passes test	
Expiration Date			31-OCT-13	
pH @ 25°C	3.99	4.01	4.00	pH
SRM Disodium Hydrogen Phosphate			186Ig	Lot
SRM Potassium Dihydrogen Orthophosphate			186Ig	Lot
SRM Potassium Hydrogen Phtalate (pH)			185h	Lot

Joe Schoellkopff

Quality Control Manager

This document has been produced electronically and is valid without a signature.

EMD Chemicals Inc.

480 South Democrat Road
Gibbstown, NJ 08027

Phone: 856-423-6300
Fax: 856-423-4389
www.emdchemicals.com

Certificate Of Analysis

Date of Release: 10.12.2011


Name: Buffer Solution pH 4.0
Buffer pH 4.00 pH $\hat{A}\pm$ 0.01, Red
NIST Traceable

Item No: BX1628, all size codes

Lot / Batch No: 1285

Country of Origin: USA

Characteristic	Requirement		Results	Units
	Min.	Max.		
Appearance			Passes test	
Expiration Date			31-OCT-13	
pH @ 25°C	3.99	4.01	4.00	pH
SRM Disodium Hydrogen Phosphate			186Ig	Lot
SRM Potassium Dihydrogen Orthophosphate			186Ig	Lot
SRM Potassium Hydrogen Phtalate (pH)			185h	Lot



Quality Control Manager

Email to Gibbstown.coa@emdchemicals.com

Astral Diagnostics, Inc.

1224 Forest Parkway
West Deptford, NJ 08066
www.astraldiagnosics.com
(800) 441-0366