

EM SCIENCE

CERTIFICATE OF ANALYSIS

NAME: Buffer Solution, pH 7
Buffer pH 7.00 ± 0.01, Yellow
ITEM NUMBER: BX1632
LOT NUMBER: 1300
FORMULA: Mixture (Phosphate)
FORMULA WT: NIST Traceable
Data Order No: 000109027

PROPERTY	LIMITS		RESULTS	UNITS
	Min.	Max.		
Appearance			Passes test	
Expiration date			31-OCT-2003	
pH at 25°C	6.99	7.01	7.005	
SRM Disodium Hydrogen Phosphate			186IIF	
SRM Potassium Dihydrogen Orthophosphate			186If	
SRM Potassium Hydrogen Phthalate (pH)			185g	



Norman Jordan,
Quality Control Manager
Release Date: 11/7/01 8:37:51 AM