

Certificate Of Analysis



Date of Release: 06.22.2011

Name: Buffer Solution, pH 10.0
Buffer pH 10.00 pH \hat{A} \pm 0.02, Blue
NIST Traceable
Item No: BX1641, all size codes
Lot / Batch No: 1157
Country of Origin: USA

Characteristic	Requirement		Results	Units
	Min.	Max.		
Appearance			Passes test	
Expiration Date			30-JUN-12	
pH @ 25 ^o C	9.98	10.02	10.00	pH
SRM Disodium Hydrogen Phosphate			186Ilg	Lot
SRM Potassium Dihydrogen Orthophosphate			186Ig	Lot
SRM Sodium Bicarbonate			191lc	Lot
SRM Sodium Carbonate			191llc	Lot

Joe Schoellkopff

Quality Control Manager

This document has been produced electronically and is valid without a signature.

EMD Chemicals Inc.

480 South Democrat Road
Gibbstown, NJ 08027

Phone: 856-423-6300
Fax: 856-423-4389
www.emdchemicals.com