

# Certificate Of Analysis



Date of Release: 12.14.2011

Name: Buffer Solution pH 3.0  
Buffer pH 3.00 pH  $\hat{A}\pm$  0.02,  
NIST Traceable  
Item No: BX1653, all size codes  
Lot / Batch No: 1347  
Country of Origin: USA

Characteristic	Requirement		Results	Units
	Min.	Max.		
Appearance			Passes test	
Expiration Date			31-DEC-13	
pH @ 25°C	2.98	3.02	3.00	
SRM Potassium Hydrogen Phtalate (pH)			185h	
SRM Potassium Tetraoxalate			189c	

Joe Schoellkopff

-----  
Quality Control Manager

This document has been produced electronically and is valid without a signature.

EMD Chemicals Inc.

480 South Democrat Road  
Gibbstown, NJ 08027

Phone: 856-423-6300  
Fax: 856-423-4389  
www.emdchemicals.com