

GFS Chemicals, Inc.
Columbus, Ohio 43223

LOT ANALYSIS

ITEM: 740 BARIUM

LOT#:P997171

Test	PASS/FAIL	NUMERICAL RESULT
Concentration 1,000 +/- 10 ug/ml	PASS	1004 ug/mL
Matrix: 1% HCl	PASS	1% HCl acid
NIST Traceable	PASS	SRM 3104a

TRACEABLE TO N.I.S.T. (Y/N)? Y

Comment:

Reported by: Kenneth L. Shafer

QC Supervisor: Charles Fontaine

Quality Assured to Retest Point: 24 months from shipment

C/A Print Date: 12/22/2015

Not for direct use in food, cosmetics, finished pharmaceuticals or drug products. Supplier is not responsible for compliance with FDA Current Good Manufacturing Practice (CGMP) requirements, including without limitation those for finished drug products in 21 C.F.R Parts 210 and 211. Consult warranty limitations at www.gfschemicals.com/statics/documents/aboutus/termsandconditions.html For resale by GFS authorized distributors only.