

GFS Chemicals, Inc.
Columbus, Ohio 43223

LOT ANALYSIS

ITEM: 8007 AMCO CLEAR TURBIDITY STANDARD, 1.0 NTU

LOT#:C033295

Test	PASS/FAIL	NUMERICAL RESULT
Turbidity 1.0 NTU +/- 0.02	PASS	1.0 NTU
NIST Traceable (avg particle size)	PASS	SRM 1963
NIST Traceable (UV-Vis and particle distribution)	PASS	SRM 2031
O.D. @ 455 nm (100 mm pathlength) 0.013-0.014	PASS	0.013
Traceable to formazin dilution value	PASS	conforms
Expiration date (1 yr from shipment)	PASS	see label

TRACEABLE TO N.I.S.T. (Y/N)? Y

Comment:

Reported by: Stephen Bailey

QC Supervisor: Rachel Edelstein

Quality Assured to Retest Point: 12 months from shipment

C/A Print Date: 12/22/2015

Not for direct use in food, cosmetics, finished pharmaceuticals or drug products. Supplier is not responsible for compliance with FDA Current Good Manufacturing Practice (CGMP) requirements, including without limitation those for finished drug products in 21 C.F.R Parts 210 and 211. Consult warranty limitations at www.gfschemicals.com/statics/documents/aboutus/termsandconditions.html For resale by GFS authorized distributors only.