

GFS Chemicals, Inc.
Columbus, Ohio 43223

LOT ANALYSIS

ITEM: 8017 AMCO CLEAR TURBIDITY STANDARD, 40 NTU

LOT#:C036600

Test	PASS/FAIL	NUMERICAL RESULT
Turbidity 40 NTU +/- 0.4	PASS	40 NTU
NIST Traceable (average particle size)	PASS	SRM 1963
NIST Traceable (particle concentration/distribution)	PASS	SRM 2031
O.D. at 455 nm (50 mm pathlength)	PASS	0.233
Traceable to fresh formazin dilution value	PASS	Conforms
Expiration date (1 year from ship date)	PASS	See label

TRACEABLE TO N.I.S.T. (Y/N)? Y

Comment:

Reported by: Renita Smith

QC Supervisor: Rachel Edelstein

Quality Assured to Retest Point: 12 months from shipment

C/A Print Date: 12/22/2015

Not for direct use in food, cosmetics, finished pharmaceuticals or drug products. Supplier is not responsible for compliance with FDA Current Good Manufacturing Practice (CGMP) requirements, including without limitation those for finished drug products in 21 C.F.R Parts 210 and 211. Consult warranty limitations at www.gfschemicals.com/statics/documents/aboutus/termsandconditions.html For resale by GFS authorized distributors only.