

1.0 ACCREDITATION / REGISTRATION

INORGANIC VENTURES is accredited to ISO Guide 34, "General Requirements for the Competence of Reference Material Producers" and ISO/IEC 17025, "General Requirements for the Competence of Testing and Calibration Laboratories".

Inorganic Ventures is also an ISO 9001 registered manufacturer (SAI Global File Number 010105).



2.0 PRODUCT DESCRIPTION

Product Code: Single Analyte Custom Grade Solution

Catalog Number: CGP10

Lot Number: J2-P02050R

Matrix: H₂O

Value / Analyte(s): 10 000 µg/mL ea:
P

Starting Material: H₃PO₄

Starting Material Lot#: 1704

Starting Material Purity: 99.9996%

3.0 CERTIFIED VALUES AND UNCERTAINTIES

Certified Value: 9990 ± 33 µg/mL - weighted mean

Certified Density: 1.015 g/mL (measured at 20 ± 1 °C)

Assay Information:

Assay Method #1 **9995 ± 40 µg/mL**
ICP Assay NIST SRM 3139a Lot Number: 060717

Assay Method #2 **9988 ± 21 µg/mL**
Acidimetric NIST SRM 84L Lot Number: 84L

- The Calculated Value is a value calculated from the weight of a starting material that has been certified directly vs. a National Institute of Standards and Technology (NIST) SRM/RM. See Sec 4.2 for balance traceability.

The following equations are used in the calculation of the certified value and the uncertainty. Reported uncertainties represent expanded uncertainties expressed at approximately the 95% confidence level using a coverage factor of k = 2.

Characterization of CRM/RM by Two Methods

Certified Value, $X_{\text{CRM/RM}}$, where two methods of characterization are used is the weighted mean of the two results:

$$X_{\text{CRM/RM}} = [(w_a)(X_a) + (w_b)(X_b)]$$

X_a = mean of Assay Method A with standard uncertainty $u_{\text{char a}}$

X_b = mean of Assay Method B with standard uncertainty $u_{\text{char b}}$

w_a and w_b = the weighting factors for each method calculated using the inverse square of the variance:

$$w_a = (1/u_{\text{char a}})^2 / ((1/u_{\text{char a}})^2 + (1/u_{\text{char b}})^2)$$

$$w_b = (1/u_{\text{char b}})^2 / ((1/u_{\text{char a}})^2 + (1/u_{\text{char b}})^2)$$

$$\text{CRM/RM Expanded Uncertainty } (\pm) = U_{\text{CRM/RM}} = k (u_{\text{char a\&b}}^2 + u_{\text{bb}}^2 + u_{\text{its}}^2 + u_{\text{sts}}^2)^{1/2}$$

k = coverage factor = 2 in all cases at Inorganic Ventures

$u_{\text{char a\&b}} = [(w_a)^2 (u_{\text{char a}})^2 + (w_b)^2 (u_{\text{char b}})^2]^{1/2}$ where $u_{\text{char a}}$ and $u_{\text{char b}}$ are the square root of the sum of the squares of errors from characterization which include instrument measurement, density, NIST SRM uncertainty, weighing, and volume

u_{bb} = bottle to bottle homogeneity standard uncertainty

u_{its} = long term stability standard uncertainty (storage)

u_{sts} = short term stability standard uncertainty (transportation)

Characterization of CRM/RM by One Method

Certified Value, $X_{\text{CRM/RM}}$, where one method of characterization is used is the mean of individual results:

$$X_{\text{CRM/RM}} = \text{mean of Assay Method A with standard uncertainty } u_{\text{char a}}$$

$$\text{CRM/RM Expanded Uncertainty } (\pm) = U_{\text{CRM/RM}} = k (u_{\text{char a}}^2 + u_{\text{bb}}^2 + u_{\text{its}}^2 + u_{\text{sts}}^2)^{1/2}$$

k = coverage factor = 2 in all cases at Inorganic Ventures

$u_{\text{char a}}$ = square root of the sum of the squares of the errors from characterization which include instrumental measurement, density, NIST SRM uncertainty, weighing, and volume

u_{bb} = bottle to bottle homogeneity standard uncertainty

u_{its} = long term stability standard uncertainty (storage)

u_{sts} = short term stability standard uncertainty (transportation)

4.0 TRACEABILITY TO NIST

- This product is traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM/RM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRM/RM are available, the term 'in-house std.' is specified.

4.1 Thermometer Calibration

- All thermometers are NIST traceable through thermometers that are calibrated by an accredited calibration laboratory.

4.2 Balance Calibration

- All analytical balances are calibrated by an accredited calibration laboratory and procedure. The weights used for testing are annually compared to master weights and are traceable to NIST.

4.3 Glassware Calibration

- An in-house procedure is used to calibrate all Class A glassware used in the manufacturing and quality control of CRM/RMs.

5.0 TRACE METALLIC IMPURITIES (TMI) DETERMINED BY ICP-MS AND ICP-OES (µg/mL)

CRM/RMs are tested for trace metallic impurities by Axial ICP-OES and ICP-MS. The result from the most sensitive method for each element, is reported below. Solutions tested by ICP-MS were analyzed in an ULPA-Filtered Clean Room. An ULPA-Filter is 99.9985% efficient for the removal of particles down to 0.3 µm.

M Ag < 0.008524	M Eu < 0.012787	O Na < 0.000100	M Se < 0.034098	O Zn < 0.000580
O Al < 0.006000	O Fe < 0.005500	M Nb < 0.002131	O Si < 0.004000	M Zr < 0.021311
M As < 0.042623	M Ga < 0.004262	M Nd < 0.008524	M Sm < 0.004262	
M Au < 0.012787	M Gd < 0.004262	M Ni < 0.026470	M Sn < 0.007411	
O B < 0.010000	M Ge < 0.025574	n Os <	M Sr < 0.002131	
M Ba < 0.042623	M Hf < 0.008524	s P <	M Ta < 0.029836	
M Be < 0.002131	O Hg < 0.012000	M Pb < 0.012787	M Tb < 0.001278	
M Bi < 0.000847	M Ho < 0.002131	M Pd < 0.021311	M Te < 0.127870	
O Ca < 0.000020	M In < 0.042623	M Pr < 0.001278	M Th < 0.004262	
M Cd < 0.003282	M Ir < 0.021311	M Pt < 0.008524	O Ti < 0.000650	
M Ce < 0.021311	O K < 0.001800	M Rb < 0.004262	M Tl < 0.004262	
M Co < 0.012787	M La < 0.002131	M Re < 0.004262	M Tm < 0.001704	
M Cr < 0.021311	O Li < 0.000020	M Rh < 0.004262	M U < 0.008524	
M Cs < 0.001278	M Lu < 0.001704	M Ru < 0.008524	M V < 0.008524	
O Cu < 0.001400	O Mg < 0.000030	O S < 0.025000	M W < 0.042623	
M Dy < 0.025574	M Mn < 0.017049	O Sb < 0.014000	M Y < 0.170494	
M Er < 0.021311	M Mo < 0.008524	M Sc < 0.042623	M Yb < 0.004262	

M - Checked by ICP-MS O - Checked by ICP-OES i - Spectral Interference

n - Not Checked For s - Solution Standard Element

6.0 INTENDED USE

- For the calibration of analytical instruments and validation of analytical methods as appropriate.

7.0 INSTRUCTIONS FOR THE CORRECT USE OF THIS REFERENCE MATERIAL

7.1 Storage and Handling Recommendations

- Store between approximately 4° - 30° C while in sealed TCT bag.
- While stored in the sealed TCT bag, transpiration of this CRM/RM is negligible. After opening the sealed TCT bag transpiration of the CRM/RM will occur, resulting in a gradual increase in the analyte concentration(s). It is the responsibility of the user to account for this effect. When the bottle is weighed both before and after being placed in storage, the mass difference observed will be a measure of transpiration mass loss.
- After opening the sealed TCT bag keep cap tightly sealed when not in use. Store and use at 20° ± 4° C. Do not pipette from the container. Do not return removed aliquots to container.
- For more information, visit www.inorganicventures.com/TCT

Atomic Weight; Valence; Coordination Number; Chemical Form in Solution - 30.97 +5 6 O=P(OH)2(O)1-

Chemical Compatibility -Soluble in HCl, HNO₃, H₂SO₄ and HF aqueous matrices water and NH₄OH. Stable with all metals and inorganic anions at low to moderate ppm levels under acidic conditions; precipitates with several metals occur in neutral media at higher concentrations.

Stability - 2-100 ppb levels- stability unknown - in 1% HNO₃ / LDPE container. 1-10,000 ppm solutions chemically stable for years in 0-1% HNO₃ / LDPE container.

P Containing Samples (Preparation and Solution) -Metal (Never found free in nature); Oxides (Water soluble); Ores (Naturally occurring only as the phosphate except for a few rare minerals found in meteorites - Na₂CO₃ fusion in Pt0); Organic Matrices (Dry ash mixed with Na₂CO₃ in Pt0 at 450EC then increase heat to 1000EC to fuse; or perform an H₂SO₄ / H₂O₂ acid digestion).

Atomic Spectroscopic Information (ICP-OES D.L.s are given as radial/axial view):

Technique/Line	Estimated D.L.	Order	Interferences (underlined indicates severe)
ICP-MS 31 amu	6000+ ppt	n/a	15N21H, 15N16O, 14N17O, 13C18O, 12C18O1H, 62Ni2+
ICP-OES 177.495 nm	0.01 / 0.005 µg/mL	1	Cu, Hf
ICP-OES 178.287 nm	0.03 / 0.002 µg/mL	1	I
ICP-OES 213.618 nm	0.08 / 0.03 µg/mL	1	Cu, Mo

8.0 HAZARDOUS INFORMATION

- Please refer to the Safety Data Sheet for information regarding this CRM/RM.

9.0 HOMOGENEITY

- This solution was mixed according to an in-house procedure and is guaranteed to be homogeneous. Homogeneity data indicate that the end user should take a minimum sample size of 0.2 mL to assure homogeneity.

10.0 QUALITY STANDARD DOCUMENTATION

10.1 10CFR50 Appendix B - Nuclear Regulatory Commission

- Domestic Licensing of Production and Utilization Facilities

10.2 10CFR21 - Nuclear Regulatory Commission

- Reporting defects and Non-Compliance

10.3 ISO 9001 Quality Management System Registration

- SAI Global File Number 010105

10.4 ISO/IEC Guide 17025 "General Requirements for the Competence of Testing and Calibration Laboratories"

- Chemical Testing - Accredited / A2LA Certificate Number 883.01

10.5 ISO/IEC Guide 34 "General Requirements for the Competence of Reference Material Producers"

- Reference Material Producer - Accredited / A2LA Certificate Number 883.02

11.0 CERTIFICATION, LOT EXPIRATION AND PERIOD OF VALIDITY

11.1 Certification Issue Date

July 22, 2015

- The certification is valid within the measurement uncertainty specified provided the CRM/RM is stored and handled in accordance with instructions given in Sec 7.1. This certification is nullified if instructions in Sec 7.1 are not followed or if the CRM/RM is damaged, contaminated, or otherwise modified.

11.2 Lot Expiration Date

- **July 22, 2018**

- The date after which this CRM/RM should not be used.

- The lot expiration date reflects the period of time that the stability of a CRM/RM can be supported by long term stability studies conducted on properly stored and handled CRM/RMs. Lot expiration is limited primarily by transpiration (loss of water from the solution) and infrequently by chemical stability.

11.3 Period of Validity

- Sealed TCT Bag Open Date: _____

- This CRM/RM should not be used longer than one year from the date of removal from the aluminized bag or after the date given in Sec. 11.2, whichever comes first. This is contingent upon the CRM/RM being handled and stored in accordance with the instructions given in Sec 7.1.

12.0 NAMES AND SIGNATURES OF CERTIFYING OFFICERS

Certificate Prepared By:

Brenda Francis
Product Documentation Technician



Certificate Approved By:

Michael Booth
QC Supervisor



Certifying Officer:

Paul Gaines
PhD., Senior Technical Director

