



Hydrochloric Acid, 0.1N Volumetric Solution

BAKER ANALYZED[®] Reagent

Product No. 5621

Lot No. L14504

Release Date 04/19/2012

Certificate of Analysis

TEST	SPECIFICATION	RESULT
Standardization at 25°C traceable to NIST Standard Reference Material.		
SRM No		723e
Normality	0.0995 - 0.1005	0.1001
Trace Impurities (in ppm):		
Sulfate (SO ₄)	1 max.	< 1
Ammonium (NH ₄)	3 max.	< 3
Heavy Metals (as Pb)	0.5 max.	< 0.3
Iron (Fe)	0.1 max.	< 0.05
For Laboratory, Research or Manufacturing Use		
USE BEFORE April 30, 2014		
Country of Origin: USA		
ISO Phillipsburg, NJ 9001.2008, 14001.2004 Paris, KY 9001.2008 Mexico City, Mexico 9001.2008 Deventer, The Netherlands 9001.2008, 14001.2004, 13485.2003 Selangor, Malaysia 9001.2008 Panaji, India 9001.2008 Gliwice, Poland 9001.2008, 17025.2005  Richard M. Siberski Global Director of Quality Assurance		

For questions on this Certificate of Analysis please contact Technical Services at 855-282-6867 or 610-573-2600

Avantor[™] Performance Materials, Inc.

3477 Corporate Parkway • Suite #200 • Center Valley, PA 18034 • U.S.A. • Phone: 610.573.2600 • Fax: 610.573.2610