



Sodium Hydroxide, 1N Volumetric Solution

BAKER ANALYZED[®] Reagent

Product No. 5635

Lot No. L14506

Release Date 04/24/2012

Certificate of Analysis

TEST	SPECIFICATION	RESULT
Standardization at 25°C traceable to NIST Standard Reference Material.		
SRM No		84L
Appearance (Clear, colorless liquid)	Passes Test	Passes Test
Normality (eq/L)	0.995 - 1.005	0.997
Trace Impurities (in ppm):		
Chloride (Cl)	5 max.	< 3
Heavy Metals (as Pb)	1 max.	< 0.5
Iron (Fe)	0.5 max.	< 0.3
Filtered through a 0.2 micron filter.		
For Laboratory, Research or Manufacturing Use		
Protect from air to avoid absorption of carbon dioxide.		
USE BEFORE April 30, 2014		
Country of Origin: USA		
<div> <div> ISO </div> <div> Phillipsburg, NJ 9001.2008, 14001.2004 Paris, KY 9001.2008 Mexico City, Mexico 9001.2008 Deventer, The Netherlands 9001.2008, 14001.2004, 13485.2003 Selangor, Malaysia 9001.2008 Panaji, India 9001.2008 Gliwice, Poland 9001.2008, 17025.2005 </div> </div> <div> <p>Richard M. Siberski Global Director of Quality Assurance</p> </div>		

For questions on this Certificate of Analysis please contact Technical Services at 855-282-6867 or 610-573-2600

Avantor[™] Performance Materials, Inc.

3477 Corporate Parkway • Suite #200 • Center Valley, PA 18034 • U.S.A. • Phone: 610.573.2600 • Fax: 610.573.2610