



Sulfuric Acid, 0.2N Volumetric Solution

BAKER ANALYZED[®] Reagent

Product No. 5690

Lot No. K20514

Release Date 06/16/2011

Certificate of Analysis

TEST	SPECIFICATION	RESULT
Standardization at 25°C traceable to NIST Standard Reference Material.		
SRM No		723D
Normality	0.1995 - 0.2005	0.2005
Trace Impurities (in ppm):		
Chloride (Cl)	1 max.	< 0.5
Phosphate (PO ₄)	1 max.	< 0.5
Heavy Metals (as Pb)	0.3 max.	< 0.2
Iron (Fe)	0.5 max.	< 0.3
For Laboratory, Research or Manufacturing Use		
USE BEFORE May 31, 2013		
Country of Origin: USA		



Phillipsburg, NJ 9001.2008, 14001.2004
Paris, KY 9001.2008
Mexico City, Mexico 9001.2008
Deventer, The Netherlands 9001.2008, 14001.2004, 13485.2003
Selangor, Malaysia 9001.2008
Panaji, India 9001.2008
Gliwice, Poland 9001.2008, 17025.2005


Richard M. Siberski
Global Director of Quality Assurance

For questions on this Certificate of Analysis please contact Technical Services at 1-800-582-2537 or 908-859-2151
Avantor[™] Performance Materials, Inc. (formerly Mallinckrodt Baker, Inc.)
222 Red School Lane • Phillipsburg, NJ 08865 • Phone: 908.859.2151 • Fax: 908.859.6905