



Calcium 1,000 ug/ml

BAKER INSTRA-ANALYZED[®] Reagent
(calcium carbonate in 5% HNO₃)



3

Product No. 6448

Lot No. L01582

Release Date 01/05/2012

Certificate of Analysis

TEST	SPECIFICATION	RESULT
Calcium(Ca)(ug/ml)	990 - 1010	1000
The standardization is traceable to NIST SRM 3109a, Calcium Standard		
Solution, Lot #		050825
Prepared by dissolution of calcium carbonate of 99.99% spectral purity. The metal content of the solution is confirmed by		
titration with EDTA, traceable to the NIST SRM 3109a.		
The balance used to weigh the raw materials is calibrated regularly using NIST traceable mass standards.		
The uncertainty associated with the reported concentration is +/-0.3%.		
For Laboratory, Research or Manufacturing Use		
USE BEFORE January 31, 2014		
Country of Origin: USA		
 Phillipsburg, NJ 9001.2008, 14001.2004 Paris, KY 9001.2008 Mexico City, Mexico 9001.2008 Deventer, The Netherlands 9001.2008, 14001.2004, 13485.2003 Selangor, Malaysia 9001.2008 Panaji, India 9001.2008 Gliwice, Poland 9001.2008, 17025.2005		
 Richard M. Siberski Global Director of Quality Assurance		

For questions on this Certificate of Analysis please contact Technical Services at 855-282-6867 or 610-573-2600

Avantor[™] Performance Materials, Inc.

3477 Corporate Parkway • Suite #200 • Center Valley, PA 18034 • U.S.A. • Phone: 610.573.2600 • Fax: 610.573.2610