



Copper, 1000 ug/mL

BAKER INTRA-ANALYZED[®] Reagent
(Cu metal in 5% HNO₃)
3

Product No. 6451
Lot No. L11603
Release Date 03/14/2012

Certificate of Analysis

TEST	SPECIFICATION	RESULT
Copper (Cu)	990 - 1010	994 ug/mL
The standardization is traceable to NIST		
SRM 915, Calcium Carbonate, Lot #		946712
Prepared by dissolution of copper metal of 99.99% spectral purity.		
The metal content of the solution is confirmed by titration with EDTA, traceable to NIST SRM 915.		
The balance used to weigh the raw materials is calibrated regularly using NIST traceable mass standards.		
The uncertainty associated with the reported concentration is +/-0.3%.		
For Laboratory, Research or Manufacturing Use		
USE BEFORE March 31, 2014		
Country of Origin: USA		



Phillipsburg, NJ 9001.2008, 14001.2004
 Paris, KY 9001.2008
 Mexico City, Mexico 9001.2008
 Deventer, The Netherlands 9001.2008, 14001.2004, 13485.2003
 Selangor, Malaysia 9001.2008
 Panaji, India 9001.2008
 Gliwice, Poland 9001.2008, 17025.2005

Richard M. Siberski
 Global Director of Quality Assurance

For questions on this Certificate of Analysis please contact Technical Services at 855-282-6867 or 610-573-2600

Avantor[™] Performance Materials, Inc.

3477 Corporate Parkway • Suite #200 • Center Valley, PA 18034 • U.S.A. • Phone: 610.573.2600 • Fax: 610.573.2610