



## 2-Propanol

**'BAKER ANALYZED'® HPLC Solvent**  
**For use in Liquid Chromatography (HPLC & UHPLC) & Spectrophotometry**

**Product No. 9095**  
**Lot No. L11E27**  
**Release Date 04/16/2012**

### Certificate of Analysis

| TEST                                         | SPECIFICATION | RESULT  |
|----------------------------------------------|---------------|---------|
| Ultraviolet Absorbance                       |               |         |
| <b>(1.00-cm cell vs. water):</b>             |               |         |
| 350 nm                                       | 0.01 max.     | < 0.002 |
| 280 nm                                       | 0.01 max.     | 0.002   |
| 254 nm                                       | 0.020 max.    | 0.005   |
| 225 nm                                       | 0.16 max.     | 0.10    |
| UV Cut-off, nm                               | 205 max.      | 203     |
| Assay (CH <sub>3</sub> CHOHCH <sub>3</sub> ) | 99.7 % min.   | 100.0 % |
| Residue after Evaporation                    | 2 ppm max.    | < 1 ppm |
| Water (by KF, coulometric)                   | 0.03 % max.   | 0.02 %  |

For Laboratory, Research or Manufacturing Use

Country of Origin: USA



Phillipsburg, NJ 9001.2008, 14001.2004  
 Paris, KY 9001.2008  
 Mexico City, Mexico 9001.2008  
 Deventer, The Netherlands 9001.2008, 14001.2004, 13485.2003  
 Selangor, Malaysia 9001.2008  
 Panaji, India 9001.2008  
 Gliwice, Poland 9001.2008, 17025.2005

*Richard M. Siberski*  
 Richard M. Siberski  
 Global Director of Quality Assurance

For questions on this Certificate of Analysis please contact Technical Services at 855-282-6867 or 610-573-2600

Avantor™ Performance Materials, Inc.

3477 Corporate Parkway • Suite #200 • Center Valley, PA 18034 • U.S.A. • Phone: 610.573.2600 • Fax: 610.573.2610