



Aniline

BAKER ANALYZED[®] Reagent

Product No. 9110

Lot No. L07N58

Release Date 03/16/2012

Certificate of Analysis

TEST	SPECIFICATION	RESULT
Assay (C ₆ H ₅ NH ₂) (by GC, corrected for water)	99.0 % min.	100.0 %
Color (APHA) (at time of packaging)	250 max.	25
Residue after Ignition	0.005 % max.	< 0.001 %
Chlorobenzene (C ₆ H ₅ Cl)	0.01 % max.	0.004 %
Hydrocarbons	Passes Test	Passes Test
Nitrobenzene (C ₆ H ₅ NO ₂)	Passes Test	Passes Test
Product Information (not specifications):		
Boiling Point (typical)	184.4 °C.	
NOTE: This reagent darkens to a reddish-brown color on storage.		
For Laboratory, Research or Manufacturing Use		
USE BEFORE December 31, 2012		
Country of Origin: USA		



Phillipsburg, NJ 9001.2008, 14001.2004
 Paris, KY 9001.2008
 Mexico City, Mexico 9001.2008
 Deventer, The Netherlands 9001.2008, 14001.2004, 13485.2003
 Selangor, Malaysia 9001.2008
 Panaji, India 9001.2008
 Gliwice, Poland 9001.2008, 17025.2005

Richard M. Siberski
 Richard M. Siberski
 Global Director of Quality Assurance

For questions on this Certificate of Analysis please contact Technical Services at 855-282-6867 or 610-573-2600

Avantor[™] Performance Materials, Inc.

3477 Corporate Parkway • Suite #200 • Center Valley, PA 18034 • U.S.A. • Phone: 610.573.2600 • Fax: 610.573.2610