

Product Specification

Bromate-Bromide Solution, 0.500 Normal (N/2), 0.250 Molar as Br₂

Lot Number: SAMPLE

Product Number: 1172

Manufacture Date: N/A

Expiration Date: N/A

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Bromide	7758-02-3	ACS
Potassium Bromate	7758-01-2	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	N/A	
Assay (vs. Potassium Iodate/Starch)	0.4998-0.5002 N at 20°C	N/A	136

Specification	Reference
Bromide-Bromate, Standard Solution (0.2500 M as Br ₂)	ASTM (D 1159)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
R1172000-10F	10 L Cubitainer®	24 months
1172-32	1 L amber glass	24 months
R1172000-20F	20 L Cubitainer®	24 months
1172-1	4 L amber glass	24 months
R1172000-4F	4 L Cubitainer®	24 months
1172-16	500 mL amber glass	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Katie Schnur
Quality Control Manager

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."