

Product Specification

Buffer Solution B, for low level Sulfate Analysis (Turbidimetric Method)

Lot Number: SAMPLE

Product Number: 1476

Manufacture Date: N/A

Expiration Date: N/A

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Magnesium Chloride Hexahydrate	7791-18-6	ACS
Sodium Acetate Trihydrate	6131-90-4	ACS
Acetic Acid	64-19-7	ACS
Potassium Nitrate	7757-79-1	ACS
Sodium Sulfate Anhydrous	7757-82-6	ACS

Test	Specification	Result
Appearance	Colorless liquid	N/A

Specification	Reference
---------------	-----------

Buffer Solution B

APHA (4500-SO42- E)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1476-32	1 L natural poly	18 months
1476-8	250 mL natural poly	18 months
1476-16	500 mL natural poly	18 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Katie Schnur
Quality Control Manager

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."