

Arlington, TX 76012
Pocomoke City, MD 21851
Batesville, IN 47006
http://www.riccachemical.com
1-888-GO-RICCA
customerservice@riccachemical.com

Product Specification

Buffer, Reference Standard, pH 6.86 ± 0.01 at 25°C

Lot Number: SAMPLE Product Number: 1540

Manufacture Date: N/A

Expiration Date: N/A

The certified value for this product is confirmed in independent testing by a second qualified chemist.

°C 10 30 35 40 50 45 6.924 6.899 6.879 6.844 Нg 6.984 6.950 6.8606.8526.8406.837 6.836

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Phosphate Dibasic	7558-79-4	ACS
Potassium Dihydrogen Phosphate	7778-77-0	ACS
Preservative	Proprietary	

Test	Specification	\mathbf{Result}	NIST SRM#
Appearance	Colorless liquid	N/A	
pH at 25°C	6.860-6.870	N/A	185 & 186
pH at 25°C Uncertainty	0.03	N/A	

Specification	Reference
Phosphate Reference Buffer Solution (nH = 6.86 at 25°C)	ASTM (D 1498)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1540-2.5	10 L Cubitainer®	24 months
1540-32	1 L natural poly	24 months
1540-5	20 L Cubitainer®	24 months
1540-20F1	20 mL pouch	24 months
1540-20	20 x 20 mL pack	24 months
1540-1	4 L natural poly	24 months
1540-1CT	4 L Cubitainer®	24 months
1540-16	500 mL natural poly	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

KatietSchnun

Katie Schnur Quality Control Manager

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference



LABORATORY
ACCREDITATION
BUREAU a division of A-S-B
ACCREDITED ISO/IEC 17022

Materials -- Contents of Certificates and Labels."

Version: 1.3 Lot Number: SAMPLE Product Number: 1540 Page 1 of 1