

Product Specification

Buffer, Reference Standard, pH 7.00 ± 0.01 at 25°C (Color Coded Green)

Lot Number: SAMPLE

Product Number: 1551.4

Manufacture Date: N/A

Expiration Date: N/A

The certified value for this product is confirmed in independent testing by a second qualified chemist.

°C	0	5	10	15	20	25	30	35	40	45	50
pH	7.12	7.09	7.06	7.04	7.02	7.00	6.99	6.98	6.98	6.97	6.97

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Phosphate Dibasic	7558-79-4	ACS
Potassium Dihydrogen Phosphate	7778-77-0	ACS
Yellow Dye	Proprietary	
Preservative	Proprietary	
Blue Dye	Proprietary	
Sodium Hydroxide	1310-73-2	Reagent

Test	Specification	Result	NIST SRM#
Appearance	Green liquid	N/A	
pH at 25°C	6.990-7.010	N/A	186 & 191
pH at 25°C Uncertainty	0.03	N/A	

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1551.4-2.5	10 L Cubitainer®	24 months
1551.4-32	1 L natural poly	24 months
1551.4-5	20 L Cubitainer®	24 months
1551.4-1	4 L natural poly	24 months
1551.4-1CT	4 L Cubitainer®	24 months
1551.4-16	500 mL natural poly	24 months
1551.4-55	55 gal drum	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Katie Schnur
Quality Control Manager

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."



LABORATORY ACCREDITATION BUREAU
a division of **A-5-B**
ACCREDITED ISO/IEC 17025