

Buffer, Reference Standard, pH 13.00 ± 0.01 at $25^{\circ}C$

Lot Number: SAMPLE

Product Number: 1625

Manufacture Date: N/A Expiration Date: N/A

The certified value for this product is confirmed in independent testing by a second qualified chemist.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Hydroxide	1310-73-2	Reagent
Potassium Chloride	7447-40-7	ACS

Test	Specification	\mathbf{Result}	NIST SRM#
Appearance	Colorless liquid	N/A	
pH at 25°C	12.990-13.010	N/A	186 & 191
pH at 25°C Uncertainty	0.03	N/A	

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with ε thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1625-2.5	10 L Cubitainer®	18 months
1625-32	1 L natural poly	18 months
1625-5	20 L Cubitainer®	18 months
1625-8	250 mL natural poly	18 months
1625-1	4 L natural poly	18 months
1625-1CT	4 L Cubitainer®	18 months
1625-16	500 mL clear PET-G	18 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

KatierSchnin

Katie Schnur Quality Control Manager

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

