

# Product Specification

**Calcium Chloride, 0.0100 Molar (M/100)**
**Lot Number: SAMPLE**
**Product Number: 1728**
**Manufacture Date: N/A**
**Expiration Date: N/A**

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Calcium Chloride Dihydrate	10035-04-8	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	N/A	
Assay (vs. EDTA/Eriochrome Black T)	0.00999-0.01001 Molar	N/A	915

Specification	Reference
Calcium chloride solution 0.01 M R	EP (1014602)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1728-32	1 L natural poly	18 months
1728-1	4 L natural poly	18 months
1728-16	500 mL natural poly	18 months

**Recommended Storage: 15°C - 30°C (59°F - 86°F)**

**Katie Schnur**  
Quality Control Manager

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."