

Product Specification

Copper Sulfate, 10% (w/v)
Lot Number: SAMPLE
Product Number: 2320
Manufacture Date: N/A
Expiration Date: N/A

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Copper Sulfate Pentahydrate	7758-99-8	ACS

Test	Specification	Result
Appearance	Blue liquid	N/A
Assay (by Refractive Index)	9.5-10.5 % (w/v)	N/A

Specification	Reference
Copper sulfate solution	TAPPI (T 430 om-94)
Copper Sulfate Solution	EPA (420.3)
Copper Sulfate Solution	EPA (420.1)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
2320-32	1 L natural poly	24 months
2320-5	20 L Cubitainer®	24 months
2320-1	4 L natural poly	24 months
2320-16	500 mL natural poly	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)


Katie Schnur
Quality Control Manager

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."