

Product Specification

DPD Indicator Solution, for Residual Chlorine Analysis

Lot Number: SAMPLE

Product Number: 2655

Manufacture Date: N/A

Expiration Date: N/A

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sulfuric Acid	7664-93-9	ACS
N,N-Diethyl-1,4-phenylenediamine Sulfate	6283-63-2	Reagent
Ethylenediaminetetraacetic Acid (EDTA), Disodium Salt, Dihydrate	6381-92-6	ACS

Test	Specification	Result
Appearance	Colorless to slightly pink liquid	N/A
Suitability for Use	Pink to red color in the presence of Chlorine	N/A

Specification	Reference
N,N-Diethyl-p-phenylenediamine (DPD) Indicator Solution	APHA (4500-CIO2 D)
N,N-Diethyl-p-phenylenediamine (DPD) Indicator Solution	APHA (4500-CI F)
N,N-Diethyl-p-phenylenediamine (DPD) Indicator Solution	APHA (4500-CI G)
N,N-Diethyl-p-phenylenediamine (DPD) Indicator Solution	EPA (330.4)
N,N-Diethyl-p-phenylenediamine (DPD) Indicator Solution	EPA (330.5)
Diethylphenylenediamine sulphate solution R	EP (1028601)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
2655-100	100 mL amber glass	9 months
2655-32	1 L amber glass	9 months
2655-1	4 L amber glass	9 months
2655-16	500 mL amber glass	9 months

Recommended Storage: 2°C - 8°C (36°F - 46°F)



Katie Schnur
Quality Control Manager

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."