

Product Specification

Drabkin's Reagent

Lot Number: SAMPLE**Product Number:** 2660**Manufacture Date:** N/A**Expiration Date:** N/A

This product is for research use only. Not for use in diagnosis or treatment of disease.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Bicarbonate	144-55-8	ACS
Potassium Ferricyanide	13746-66-2	ACS
Potassium Cyanide	151-50-8	ACS

Test	Specification	Result
Appearance	Yellow-green liquid	N/A

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
2660-32	1 L amber poly	24 months
2660-16	500 mL amber poly	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)Katie Schnur
Quality Control Manager

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."