

# Product Specification

**Aerosol OT, 10% (w/v)****Lot Number: SAMPLE****Product Number: 281****Manufacture Date: N/A****Expiration Date: N/A**

Reagent Alcohol consists of approximately 90% Ethyl Alcohol (CH<sub>3</sub>CH<sub>2</sub>OH, CAS No. 64-17-5), 5% Methyl Alcohol (CH<sub>3</sub>OH, CAS No. 67-56-1), and 5% Isopropyl Alcohol (CH<sub>3</sub>CHOHCH<sub>3</sub>, CAS No. 67-63-0).

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Ethyl Alcohol	64-17-5	ACS
Docosate Sodium	577-11-7	Purified
Methyl Alcohol	67-56-1	ACS
Isopropyl Alcohol	67-63-0	ACS

Test	Specification	Result
Appearance	Colorless to slightly yellow liquid	N/A

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
281-32	1 L natural poly	18 months
281-1	4 L natural poly	18 months
281-16	500 mL natural poly	18 months

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)Katie Schnur  
Quality Control Manager

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."