

# Product Specification

## Eosin Y TS, 0.5% (w/v), Adsorption Indicator for Argentometric Titrations

**Lot Number:** SAMPLE

**Product Number:** 2858

**Manufacture Date:** N/A

**Expiration Date:** N/A

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Eosin Y	17372-87-1	Certified

Test	Specification	Result
Appearance	Deep red liquid	N/A
Suitability for Use	Magenta (excess Silver ions)	N/A

Specification	Reference
Eosin Y Reagent Solution	ACS (N/A)
Eosin Y TS	USP (N/A)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
2858-4	120 mL natural poly	24 months
2858-16	500 mL natural poly	24 months

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)



Katie Schnur  
Quality Control Manager

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."