

Product Specification

Fehling's Copper Solution (Fehling Solution A) for Determination of Reducing Sugars

Lot Number: SAMPLE

Product Number: 3010

Manufacture Date: N/A

Expiration Date: N/A

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Copper Sulfate Pentahydrate	7758-99-8	ACS

Test	Specification	Result	NIST SRM#
Appearance	Blue liquid	N/A	
Assay (vs. Sodium Thiosulfate/Starch)	0.2770-0.2780 N at 20°C	N/A	136

Specification	Reference
Cupri-tartaric solution R Solution I	EP (1023300 Solution I)
Cupric Tartrate TS, Alkaline (Fehling's Solution) – The Copper Solution (A)	USP (N/A)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with ϵ thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
3010-2.5	10 L Cubitainer®	24 months
3010-32	1 L natural poly	24 months
3010-5	20 L Cubitainer®	24 months
3010-1	4 L natural poly	24 months
3010-16	500 mL natural poly	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Katie Schnur
Quality Control Manager

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."