

Product Specification

Formaldehyde, 10% (v/v) (non-buffered), Histological Fixative

Lot Number: SAMPLE**Product Number:** 3180**Manufacture Date:** N/A**Expiration Date:** N/A

For In Vitro Diagnostic Use. Suitable for use as a histological fixative for stain preparation and preservation of tissue samples. May be suitable for other laboratory applications. Consult standard text and operating procedures for detailed instructions.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Formaldehyde	50-00-0	ACS
Methyl Alcohol	67-56-1	ACS

Test	Specification	Result
Appearance	Colorless liquid	N/A

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
3180-2.5	10 L Cubitainer®	18 months
3180-32	1 L natural poly	18 months
3180-5	20 L Cubitainer®	18 months
3180-1	4 L natural poly	18 months
3180-16	500 mL natural poly	18 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Katie Schnur
Quality Control Manager

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."