

Product Specification

Hydrochloric Acid, 1.00 Normal

Lot Number: SAMPLE

Product Number: 3700

Manufacture Date: N/A

Expiration Date: N/A

The certified value for this product is confirmed in independent testing by a second qualified chemist.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Hydrochloric Acid	7647-01-0	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	N/A	
Assay (vs. Tris(hydroxymethyl)aminomethane/Methyl Red)	0.999-1.001 N at 20°C	N/A	723

Specification	Reference
Hydrochloric Acid, 1 N	APHA (3500-Se E)
Hydrochloric Acid Solution, 1 N	APHA (4500-NO3- B)
Hydrochloric acid, 1N HCl	TAPPI (T 613 cm-83)
Hydrochloric Acid (1 N)	EPA (SW-846) (1311)
Hydrochloric acid, approximately 1 N	TAPPI (T 255 cm-97)
Hydrochloric acid, 1 M	TAPPI (T 408 cm-97)
Hydrochloric acid, 1 N	TAPPI (T 610 sp-97)
Hydrochloric Acid, 1 N	ASTM (D 4047)
1 N hydrochloric acid	TAPPI (T 553 pm-92)
Hydrochloric Acid, Standard Solution (1.0 N)	ASTM (D 2187 I)
Hydrochloric Acid, 1 Normal Volumetric Solution	ACS (N/A)
1 M Hydrochloric acid Volumetric Solution	EP (3001800)
Hydrochloric Acid, Normal (1 N) Volumetric Solution	USP (N/A)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
3700-2.5	10 L Cubitainer®	24 months
3700-4	120 mL natural poly	24 months
3700-32CS	12 x 1L natural poly	24 months
3700-16CS	12 x 500 mL natural poly	24 months
3700-32	1 L natural poly	24 months
3700-5HP	20 L Ropak™	24 months
3700-5	20 L Cubitainer®	24 months
3700-30	30 gal drum	24 months
3700-1	4 L natural poly	24 months
3700-1CT	4 L Cubitainer®	24 months

3700-1CS	4 x 4 L natural poly	24 months
3700-16	500 mL natural poly	24 months
3700-55	55 gal drum	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Katie Schnur
Quality Control Manager

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."