

Product Specification

Hydrochloric Acid, 6.00 Normal

Lot Number: SAMPLE

Product Number: 3750

Manufacture Date: N/A

Expiration Date: N/A

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Hydrochloric Acid	7647-01-0	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless to slightly greenish-yellow liquid	N/A	
Assay (vs. Sodium Hydroxide/Phenolphthalein)	5.995-6.005 N at 20°C	N/A	84

Specification	Reference
Hydrochloric Acid, 6 N	APHA (4500-S2- F)
Hydrochloric Acid, 6 N	APHA (4500-S2- E)
Hydrochloric Acid, 6 N	APHA (2530 C)
Hydrochloric Acid, 6 N	APHA (4500-NO3- E)
Hydrochloric acid, 6 M	TAPPI (T 244 om-93)
Hydrochloric acid, 6M	TAPPI (T 266 om-94)
Hydrochloric Acid, 6 N	EPA (376.1)
Hydrochloric Acid (6 N)	EPA (SW-846) (9030B)
Hydrochloric acid, 6 M	TAPPI (T 247 cm-83)
Dilute Hydrochloric Acid, 6 N	EPA (353.3)
Dilute Hydrochloric Acid, 6 N	EPA (353.2)
6 M Hydrochloric acid Volumetric Solution	EP (3001500)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
3750-2.5	10 L Cubitainer®	24 months
3750-32	1 L natural poly	24 months
3750-5	20 L Cubitainer®	24 months
3750-8	250 mL natural poly	24 months
3750-25A	25 mL poly dropper	24 months
3750-30A	30 mL natural poly	24 months
3750-40A	40 mL natural poly	24 months
3750-1	4 L natural poly	24 months
3750-1CT	4 L Cubitainer®	24 months
3750-16	500 mL natural poly	24 months
3750-55	55 gal drum	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

A handwritten signature in black ink that reads "Katie Schnur". The signature is written in a cursive style and is placed on a light yellow rectangular background.

Katie Schnur
Quality Control Manager

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."