

# Product Specification

## Phosphate Buffer, pH 7.2, for Biochemical Oxygen Demand (BOD)

**Lot Number:** SAMPLE

**Product Number:** 5800

**Manufacture Date:** N/A

**Expiration Date:** N/A

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Phosphate, Dibasic	7758-11-4	ACS
Ammonium Chloride	12125-02-9	ACS
Sodium Phosphate Dibasic	7558-79-4	ACS
Phosphoric Acid	7664-38-2	ACS
Potassium Dihydrogen Phosphate	7778-77-0	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	N/A	
pH at 25°C	7.15-7.25	N/A	186 & 191

Specification	Reference
Phosphate Buffer Solution	APHA (5210 B)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
5800-32	1 L natural poly	24 months
5800-1	4 L natural poly	24 months
5800-16	500 mL natural poly	24 months

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)



Katie Schnur  
Quality Control Manager

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."