

# Product Specification

## Potassium Chloride Conductivity Standard, 50 µS/cm at 25°C

**Lot Number:** SAMPLE

**Product Number:** 5886.5

**Manufacture Date:** N/A

**Expiration Date:** N/A

The certified value for this product is confirmed in independent testing by a second qualified chemist.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Chloride	7447-40-7	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	N/A	
Conductivity at 25°C	49.9-50.1 µS/cm	N/A	3193
Conductivity at 25°C Uncertainty	0.32	N/A	

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
5886.5-2.5	10 L Cubitainer®	12 months
5886.5-32	1 L natural poly	12 months
5886.5-5	20 L Cubitainer®	12 months
5886.5-8	250 mL natural poly	12 months
5886.5-1	4 L natural poly	12 months
5886.5-16	500 mL natural poly	12 months

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)



Katie Schnur  
Quality Control Manager

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."


**LABORATORY  
ACCREDITATION  
BUREAU** a division of **A-S-B**  
**ACCREDITED** ISO/IEC 17025