Arlington, TX 76012
Pocomoke City, MD 21851
Batesville, IN 47006
http://www.riccachemical.com
1-888-GO-RICCA
customerservice@riccachemical.com

## **Product Specification**

Potassium Dichromate, 0.250 Normal (N/4), 0.04167 Molar (M/24)

Lot Number: SAMPLE Product Number: 6070

Manufacture Date: N/A

Expiration Date: N/A

Name	CAS#	Grade	
Water	7732-18-5	ACS/ASTM/USP/EP	
Potassium Dichromate	7778-50-9	ACS	

Test	Specification	$\mathbf{Result}$	NIST SRM#
Appearance	Orange liquid	N/A	
Assay (vs. Sodium Thiosulfate/Starch)	0.2499-0.2501 N at 20°C	N/A	136

Specification	Reference
Potassium Dichromate Solution, Standard (0.25 N)	ASTM (D 1252 A)
Standard Potassium Dichromate Solution, 0.04167 M	APHA (5220 B)
Standard Potassium Dichromate Solution, 0.25 N	EPA (410.3)
Standard Potassium Dichromate Solution, 0.250 N	EPA (410.1)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with \$\epsilon\$ thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
6070-2.5	10 L Cubitainer®	24 months
6070-32	1 L natural poly	24 months
6070-5	20 L Cubitainer®	24 months
6070-1	4 L natural poly	24 months
6070-16	500 mL natural poly	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

KatietSchnun

Katie Schnur

Quality Control Manager

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

Version: 1.3 Lot Number: SAMPLE Product Number: 6070 Page 1 of 1