

# Product Specification

**Sodium Hydroxide, 5% (w/v) (50 g/L)**
**Lot Number: SAMPLE**
**Product Number: 7250**
**Manufacture Date: N/A**
**Expiration Date: N/A**

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Hydroxide	1310-73-2	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	N/A	
Assay (vs. Sulfuric Acid/Phenolphthalein)	4.96-5.04 % (w/v)	N/A	723

Specification	Reference
Sodium Hydroxide Solution	APHA (413 B)
Sodium Hydroxide Solution	EPA (335.1)
Sodium Hydroxide Solution (1.25 N)	EPA (SW-846) (9010B)
Sodium Hydroxide Solution (1.25 N)	EPA (SW-846) (9012A)
Sodium Hydroxide Solution (1.25 N)	EPA (SW-846) (7.3.3.2)
Sodium Hydroxide Solution (1.25 N)	EPA (SW-846) (7.3.4.2)
Sodium hydroxide (1.25 N)	EPA (SW-846) (9010B)
Sodium hydroxide (1.25 N)	EPA (SW-846) (9012A)
Sodium Hydroxide Solution	ASTM (D 4046)
Sodium Hydroxide Solution, 1.25 N	EPA (335.2)
Sodium Hydroxide Solution (50 g/L)	ASTM (D 1126)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
7250-10A	10 mL natural poly	36 months
7250-4	120 mL natural poly	36 months
7250-32	1 L natural poly	36 months
7250-5	20 L Cubitainer®	36 months
7250-8	250 mL natural poly	36 months
7250-1	4 L natural poly	36 months
7250-16	500 mL natural poly	36 months
7250-55	55 gal drum	36 months

**Recommended Storage: 15°C - 30°C (59°F - 86°F)**

A handwritten signature in black ink that reads "Katie Schnur". The signature is written in a cursive style and is placed on a light yellow rectangular background.

Katie Schnur  
Quality Control Manager

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."