

# Product Specification

**Arsenic AA Standard, 1000 ppm As in 3% HNO<sub>3</sub>**
**Lot Number: SAMPLE**
**Product Number: AAS1KN**
**Manufacture Date: N/A**
**Expiration Date: N/A**

This is a single element solution that was prepared volumetrically to contain the certified value reported. The uncertainty associated with the certified value is the sum of the estimated errors due to the purity of the raw material, the volumetric preparation of the solution, and transpiration of the solution through the container wall.

The final solution concentration is confirmed by AA, ICP, or ICP-MS, and is traceable to NIST Standard Reference Material 3103.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Nitric Acid	7697-37-2	Trace Metals
Arsenic Trioxide	1327-53-3	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	N/A	
Arsenic (As)	995-1005 ppm	N/A	3103

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
AAS1KN-100	100 mL natural LDPE	24 months
AAS1KN-500	500 mL natural poly	24 months

**Recommended Storage: 15°C - 30°C (59°F - 86°F)**


Katie Schnur  
Quality Control Manager

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."