

Product Specification

Acetone/Acetonitrile, 80/20, HPLC Grade

Lot Number: SAMPLE

Product Number: R0195000

Manufacture Date: N/A

Expiration Date: N/A

Name	CAS#	Grade
Acetone	67-64-1	HPLC
Acetonitrile	75-05-8	HPLC

Test	Specification	Result
Appearance	Colorless liquid	N/A

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
R0195000-20E4	20 L metal can	24 months
R0195000-4C	4 L amber glass	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Katie Schnur
Quality Control Manager

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."