

Product Specification

Potassium Chloride Conductivity Standard, 1413 $\mu\text{S/cm}$ at 25°C

Lot Number: SAMPLE

Product Number: R5888141

Manufacture Date: N/A

Expiration Date: N/A

The certified value for this product is confirmed in independent testing by a second qualified chemist.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Chloride	7447-40-7	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	N/A	
Conductivity at 25°C	1412-1414 $\mu\text{S/cm}$	N/A	3193
Conductivity at 25°C Uncertainty	5.8	N/A	

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
R5888141-10F	10 L Cubitainer®	24 months
R5888141-1A	1 L natural poly	24 months
R5888141-20F	20 L Cubitainer®	24 months
R5888141-4A	4 L natural poly	24 months
R5888141-500A	500 mL natural poly	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Katie Schnur
Quality Control Manager

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."


LABORATORY ACCREDITATION BUREAU
a division of A-S-B
ACCREDITED ISO/IEC 17025